



1. PERFORMANCE


- 1) Measuring range : 0.05-2.5 %
- Number of pump strokes : 2 (200mℓ)
- 2) Sampling time : 3 minutes/2 pump strokes
- 3) Detectable limit : 10 ppm
- 4) Shelf life : 3 years
- 5) Operating temperature : 0 ~ 40 °C
- 6) Reading : Graduations printed on the tube are calibrated by Methyl ethyl ketone at 2 pump strokes and 1,4-Dioxane concentration is determined by using a conversion chart.
- 7) Colour change : Orange → Brownish green

2. RELATIVE STANDARD DEVIATION

RSD-low : 10% RSD-mid. : 5% RSD-high : 5%

3. CHEMICAL REACTION

Dichromate is reduced.




4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	ppm	Coexistence
Acetylene		3%	Whole reagent is discoloured to Brown.
Propane		0.2%	“
Other organic or vapours except Halogenated hydrocarbons	Similar stain is produced.	50	Higher reading are given.

Dioxane (%)



No.139SB Tube reading (%)